



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/768,565	01/29/2004	Karla Weaver	10123/00801	6338
7590 07/02/2007				
Patrick J. Fay, Esq. FAY KAPLUN & MARCIN, LLP Suite 702 150 Broadway New York, NY 10038				
			EXAMINER	
			STIGELL, THEODORE J	
			ART UNIT	PAPER NUMBER
			3763	
			MAIL DATE	DELIVERY MODE
			07/02/2007	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/768,565

Applicant(s)

WEAVER ET AL.

Examiner

Theodore J. Stigell

Art Unit

3763

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 December 2006.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-21 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-21 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- ☒ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☒ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date 11/06/2006
- ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- ☐ Notice of Informal Patent Application
- ☐ Other: _____

DETAILED ACTION

Response to Amendment

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1-21 are rejected under 35 U.S.C. 102(b) as being anticipated by Mikhail et al. ('357).

Mikhail et al. discloses a catheter comprising a flow control membrane (38). The flow control membrane comprises a plurality of slits (76). The slits are substantially symmetrical about either a line of symmetry, diameter, major axis or a point. (See Figures 21-35) The flow control membrane is composed of silicone rubber. (See Column 18 Lines 46-58) The flow control membrane is taught to be shapes other than circular such as oval. (See Column 27 Lines 64-65)

Response to Arguments

Applicant's arguments filed 12/06/2006 have been fully considered but they are not persuasive.

In response to the Applicants' argument that Mikhail does not disclose a "pressure actuated valve including a flow control membrane... including a plurality of slits extending therethrough, wherein, when the membrane is acted upon by a pressure of at least a threshold magnitude, the slits open to permit flow", the Examiner

respectfully disagrees. The Applicant argues that the device of Mikhail is designed to be opened by a squeezing force and is not opened by fluid pressure. However, there is no limitation in the claim that requires the valve to be opened by fluid pressure.

Furthermore, even if the fluid pressure were claimed, the device of Mikhail would still meet this functional limitation. The Applicants' contention that "the valve of Mikhail is designed to remain sealed at all times regardless of the pressure applied thereto" is incorrect. While the valve is designed to resist normal pressures associated with urine drainage, there is obviously a flow pressure at which the valve would open.

In response to the Applicant's argument that Mikhail does not disclose a plurality of slits with each of the slits extending "between end portions thereof along a curve and wherein a distance between a first end portion of a first one of the slits and a first end portion of a second one of the slits is a minimum distance between the first and second slits", the Examiner respectfully disagrees. The Applicant has merely stated this limitation is not anticipated and has not provided any substantial arguments to support this statement. The Examiner contends that this limitation is anticipated.

In response to the Applicant's argument that Mikhail does not disclose a "substantially planar elastic membrane", the Examiner respectfully disagrees. The term planar simply means "lying in a plane". The valve (38) certainly lies in a plane.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.


Any inquiry concerning this communication or earlier communications from the examiner should be directed to Theodore J. Stigell whose telephone number is 571-272-8759. The examiner can normally be reached on M-F 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Nicholas Lucchesi can be reached on 571-272-4977. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3763

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.


Theodore J. Stigell


NICHOLAS D. LUCCHESI
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 3700